

Declaration of Conformity

EEC guidelines 93/42/EEC Appendix VII

max-c

side 01

Date: 01.08.2005

We herewith declare that the electric drive for weelchairs

max-e

has been developed and produced in concordance with the EEC guidelines 93/42/EEC.

This device belongs into category I of medical products according to EEC guidelines. The CE identification confirms that the device corresponds to these guidelines. A technical documentation plus risk analysis is avilable.

The following norms and specifications have been applied:

Mechanical safety:

DIN EN 12184

As well as the norms:

DIN EN 12182 ISO 7176- 1 bis 10 ISO 7176-15 DIN EN 1041

Electrical safety:

DIN EN 12184 ISO 7176-14

As well as the norms:

EN 60601-1 und IEC 60513

Electromagnetic

tolerance:

DIN EN 12184

ISO 7176-21

The product's user's manual is available:

- a) in the original version
- b) in the user's mother tongue

The management of the company AAT Alber-Antriebstechnik GmbH,

Ehestetter Weg 11 72458 Albstadt (Germany).

Thomas Alber, manager

01.08.2005

Markus Alber, manager

01.08.2005